

RFP NO. 6417/03/12/2024: Request for Quotation (RFQ) for the supply of Tedlar bags sampling pumps, Tedlar bags, tubings and fittings to the CSIR

RESPONSE TO QUESTIONS

Please see response below:

Please would you provide clarify on he below re: tender as mentioned in my subject line:

1. As per your line 2: 1X Light weight dust and vapour grab sampling pump (include operater manuals) with fittings for use on Tedlar bags. Qty 1, the following questions:
 1. What substances will you be sampling and subsequently analysing for? **HF, HSiF6, SO2,**
 2. What flow rate is required for this pump? **We prefer to follow the below standards:**

Pump Type	Sample time and/or volume for 5L bag and 33,567 mls volume			
	Grab sample	1 hour	8 hours	24 hours
AENV (10-1500 ml/min)	15 min @ 1.5 l/min	450 ml/min	55 ml/min	20 ml/min
SKC (1-5000 ml/min)	5 min @ 5 l/min	450 ml/min	55 ml/min	20 ml/min

2. As pe your line 3: 20 times 1Liter capacity Tedlar bags with fitting connectors for four (4 months). Qty 2(PK/10), the following:
 - What do you mean by four (4 Months)? **20 tedlar bags per month for 4 months**